

RT-PCT Test Quality Control

—Chen Yu Hui BGI Research


- Prior to sample handling
- Check points during experiment
- Quality control of test data

Sample Collection — Sample Pretreatment — RNA Extraction — qPCR Reaction — Result Calling

Prior to sample handling

- Laboratory layout and designated function of each area
- Equipment Maintenance
- Reagent and Consumables QC

Prior to sample handling - Laboratory layout







***Designated Function *Unidirectional Workflow *Proper pressure to contain nucleic acid in the expected area**

Prior to sample handling - Maintenance; Consumables QC



$$Q_{\text{Total volume}} = Q_{\text{Downflow}} + Q_{\text{Inflow}}$$
$$Q_{\text{Air exhaust}} = Q_{\text{Inflow}}$$



#Function #Calibration #Proper Usage Training

Prior to sample handling – Reagent QC & Verification

Key reagent: Nucleic Acid Extraction Kit and RT-PCT Kit



#Nucleic Acid Extraction Kit: Concentration and Purity

- Nanodrop or Qubit:A260/A280 & Electrophoresis
- Ct value: Control samples with known concentration

#RT-PCT Kit

- Sensitivity & Specificity
- Repeatability & Limit of Detection (LOD)

Conc.	Channel	1	2	3	4
10000copies/ml	FAM				
	VIC				
5000copies/ml	FAM				
	VIC				
2500copies/ml	FAM				
	VIC				




***It is suggested, at least, to have two types of RT-PCT kit for result verification**

Check points during experiment

- Sample arrival status & Sample registration
- Sample assignment into consumables
- RNA extraction & RT-PCR reaction

Check points - Sample arrival status & Sample registration

❖ Samples should reach the lab in the box at **DESIABLE TEMPERATURE**




↑ ↑ ↑ ↑
Provide Sample No. and Name of each sample in excel sheet to the lab BEFORE ship out the samples

HARD COPY of Information sheet with **SCANNABLE** barcode on the sheet is acceptable.

Sample Information/barcode provided in Hand-writing form will end up in sample rejection!!!


Check points-Sample arrival status & Sample registration



ONE TUBE, ONE BAG

*Scannable barcode

*Barcode paper should be *heat-resistance*




**One sample leaks,
All samples will be rejected**

Check points - Sample assignment

COVID-19检测任务单/ COVID-19 Sample Nucleic Acid Test Record												A1	KA0004817	
板号/Plate ID: 20200407-1 排单人/Arrange Staff: HZB 排单时间/Time:						提取试剂配液编号/Extraction Reagent Batch ID: 0318-E1						B1	KA0003576	
Sample ID	1	2	3	4	5	6	7	8	9	10	11	12	C1	KA0004837
A	KA0004817	KA0003944	KA0004804	KA0004845									D1	KA0004827
	1 <input type="checkbox"/>	9 <input type="checkbox"/>	17 <input type="checkbox"/>	25 <input type="checkbox"/>	33 <input type="checkbox"/>	41 <input type="checkbox"/>	49 <input type="checkbox"/>	57 <input type="checkbox"/>	65 <input type="checkbox"/>	73 <input type="checkbox"/>	81 <input type="checkbox"/>	89 <input type="checkbox"/>	E1	KA0004810
B	KA0003576	KA0003572	KA0004822	KA0004818									F1	KA0003578
	2 <input type="checkbox"/>	10 <input type="checkbox"/>	18 <input type="checkbox"/>	26 <input type="checkbox"/>	34 <input type="checkbox"/>	42 <input type="checkbox"/>	50 <input type="checkbox"/>	58 <input type="checkbox"/>	66 <input type="checkbox"/>	74 <input type="checkbox"/>	82 <input type="checkbox"/>	90 <input type="checkbox"/>	G1	KA0004833
C	KA0004837	KA0003583	Blank	KA0003591									H1	KA0003586
	3 <input type="checkbox"/>	11 <input type="checkbox"/>	19 <input type="checkbox"/>	27 <input type="checkbox"/>	35 <input type="checkbox"/>	43 <input type="checkbox"/>	51 <input type="checkbox"/>	59 <input type="checkbox"/>	67 <input type="checkbox"/>	75 <input type="checkbox"/>	83 <input type="checkbox"/>	91 <input type="checkbox"/>	A2	KA0003944
D	KA0004827	KA0003583	KA0004845	KA0003210									B2	KA0003572
	4 <input type="checkbox"/>	12 <input type="checkbox"/>	20 <input type="checkbox"/>	28 <input type="checkbox"/>	36 <input type="checkbox"/>	44 <input type="checkbox"/>	52 <input type="checkbox"/>	60 <input type="checkbox"/>	68 <input type="checkbox"/>	76 <input type="checkbox"/>	84 <input type="checkbox"/>	92 <input type="checkbox"/>	C2	KA0003585
E	KA0004810	KA0003574	KA0004842										D2	KA0003583
	5 <input type="checkbox"/>	13 <input type="checkbox"/>	21 <input type="checkbox"/>	29 <input type="checkbox"/>	37 <input type="checkbox"/>	45 <input type="checkbox"/>	53 <input type="checkbox"/>	61 <input type="checkbox"/>	69 <input type="checkbox"/>	77 <input type="checkbox"/>	85 <input type="checkbox"/>	93 <input type="checkbox"/>	E2	KA0003574
F	KA0003578	KA0004844	KA0004806										F2	KA0004844
	6 <input type="checkbox"/>	14 <input type="checkbox"/>	22 <input type="checkbox"/>	30 <input type="checkbox"/>	38 <input type="checkbox"/>	46 <input type="checkbox"/>	54 <input type="checkbox"/>	62 <input type="checkbox"/>	70 <input type="checkbox"/>	78 <input type="checkbox"/>	86 <input type="checkbox"/>	94 <input type="checkbox"/>	G2	KA0003597
G	KA0004833	KA0003597	KA0004840										H2	KA0003214
	7 <input type="checkbox"/>	15 <input type="checkbox"/>	23 <input type="checkbox"/>	31 <input type="checkbox"/>	39 <input type="checkbox"/>	47 <input type="checkbox"/>	55 <input type="checkbox"/>	63 <input type="checkbox"/>	71 <input type="checkbox"/>	79 <input type="checkbox"/>	87 <input type="checkbox"/>	95 <input type="checkbox"/>	A3	KA0004804
H	KA0003586	KA0003214	KA0004830										B3	KA0004822
	8 <input type="checkbox"/>	16 <input type="checkbox"/>	24 <input type="checkbox"/>	32 <input type="checkbox"/>	40 <input type="checkbox"/>	48 <input type="checkbox"/>	56 <input type="checkbox"/>	64 <input type="checkbox"/>	72 <input type="checkbox"/>	80 <input type="checkbox"/>	88 <input type="checkbox"/>	96 <input type="checkbox"/>	C3	Blank
备注/Remark:						qPCR试剂配液编号/qPCR Reagent Mix Batch ID: 0318-Q1						D3	KA0004843	
取样操作人/Sampling Staff:			复核人/Verifier:			取样时间/Sampling Time:						E3	KA0004842	
提取操作人/Nucleic Acid Extraction Staff:			复核人/Verifier:			提取仪器编号/Nucleic Acid Extraction MGISP-960 No.: 6						F3	KA0004806	
上机操作人/qPCR Operator:			qPCR仪编号/qPCR Machine No.:			上机时间/qPCR Experiment Time:						G3	KA0004840	


Sample layout for RNA extraction




Sample loading verified by two technicians

Check points - RNA extraction & RT-PCR reaction


Sample volume matters!



Miss adding of reagent or sample



NO Signal Captured




Unusual Amplification Plot

Evaporation


Quality control of test data - Key Concepts

- **Baseline:** The signal level during the initial cycles of PCR, usually cycles 3 to 15, in which there is little change in fluorescent signal.
- **Threshold:** A statistically significant increase over the calculated baseline signal. Usually sets the threshold at **10 times the standard deviation** of the fluorescence value of the baseline.
- **Ct:** Cycle number at which the **fluorescent signal of the reaction crosses the threshold. So no cross, no Ct.**




Quality control of test data – Control samples

- Positive Control: To check whether Viral RNA can be detected
- Blank control: To monitor the contamination level or background noise level
- Negative control: To check whether beta-actin gene can be amplified successfully without Viral RNA signal




- ❖ Both viral RNA and beta-actin DNA/RNA should be extracted and then amplified during qPCR reaction, which realizes the quality control of nucleic acid extraction. No beta-actin gene amplification or unusual amplification plot will lead to retest.

Quality control of test data – Example of result with Ct value



Quality control of test data – Example of result with Ct value




Why non-specific amplification in blank control?

- 1) Primer design is not perfect causing primer dimer formation leading to amplification in the end of PCR cycles;
- 2) Dye in the reagent kit partly degrades evidently in the end of cycles;
- 3) Primer in the kit is too concentrated, thus generating non-specific peaks for blank controls;
- 4) Slight contamination during sample mixing;
- 5) Lab environment is not clean enough with accumulated DNA/RNA aerosol;
- 6) 7500 is not well calibrated;

.....

- Always, *repeating the test and * thoroughly clean the laboratory area and related equipment will alleviate this situation. Separating RNA extraction and reagent mixing in different Bio-safety cabinets is suggested.
- Manually move the threshold to cover the noise and reanalyse Ct value for the rest of sample is suggested before data interpretation.

Threshold moved up



THANK YOU!